

Nova Star Mid Range DNA Ladder

Reference: AB15015



1 of 2

INTENDED USE AND PRESENTATION:

This ladder is widely used in molecular biology research.

AB15015, 50 µg. Concentration of 0,1 µg/µL.

For research use only.

SUMMARY, EXPLANATION AND LIMITATIONS:

Nova Star Mid Range DNA Ladder can be used for precise visual sizing of double stranded DNA fragments from 300 to 5000 bp on agarose gels. Supplied in sufficient quantity for 100 loadings using 4-5 µl per loading.

For best results, please load 5 µl of the Nova Star Mid Range DNA Ladder per well. (Do not forget to premix with appropriate loading dye.)

Note: Not designed for quantitating DNA concentration in a sample.




Figure 1. Nova Star Mid Range DNA Ladder showing DNA fragments from 300-5000.

APPLICATIONS:

Nova Star Mid Range DNA Ladder is a ladder marker for size determination of PCR generated DNA fragments. This marker can be used with either agarose or polyacrylamide gels. The recommended agarose gel concentration is 1,5% for this marker.

PRODUCT COMPOSITION:

Consists of ten fragments: 300 bp, 500 bp, 500 bp, 1000 bp, 1500 bp, 2000 bp, 2500 bp, 3000 bp, 4000 bp and 5000 bp.

METHODS AND PROCEDURE:

This ladder needs to be mixed with loading dye of choice using 4-5 µL per loading.

REQUIRED MATERIALS BUT NOT SUPPLIED:

All reagents, materials, and laboratory equipment for PCR and determination procedures are not provided with this ladder. This includes sterile reaction tubes, micropipettes and tips, template DNA, gen-specific PCR primer pair, dNTPs mixture, PCR grade H₂O, heat pretreatment equipment (thermoblock, microwave), centrifuge, cold store and thermal block cycler.

Buffered solutions for DNA extraction or purification, enzyme treatments, highly sensitive detection systems, and other auxiliary reagents are available from Genova Scientific.

STORAGE AND STABILITY:

Store at -20°C until the expiration date printed on product label. Avoid prolonged exposure to light. Avoid multiple freeze-thaw cycles and exposure to frequent temperature changes. Do not use after the expiration date. If the product is stored under different conditions from those stipulated in these technical indications, the new conditions must be verified by the user. The validity period of the ready to use products when opened, is the same as the expiration date indicated on the label of intact product.

Genova Scientific guarantees that the product will maintain all of the described characteristics from the production date until the expiration date, as long as the product is stored and used as recommended. No other guarantees are provided. Under no circumstances Genova Scientific is obliged to cover damages caused by use of this reagent.

TROUBLESHOOTING:

If unusual banding is observed or any other deviations from the expected results, please read these instructions carefully, along with the instructions from the PCR and determination systems. If this does not solve the problem, please contact Genova Scientific's technical support department (techsupport@genovalab.com) or your local distributor.

PRECAUTIONS:

Use only by qualified personnel.

Use proper protective equipment in order to avoid contact with reagents and samples in the eyes, skin, and mucosal tissues. In case of contact with sensitive areas, immediately flush the affected area with water. Avoid microbial contamination of the reagent, as this may produce nonspecific amplification results.

Material safety data sheet (MSDS) is available upon request.

PERFORMANCE CHARACTERISTICS:

Genova Scientific has performed studies to evaluate the functioning of this ladder for use with standard visualization and determination systems, concluding that the product is both specific and sensitive for determination performance.



Catalog number



Batch code



Research use only



Temperature limitation



Expiration date



Manufacturer



See instruction for use



Genova Scientific, S.L.
C/ Johann Gutenberg, 4F. Pol. Ind.
El Cafamo I • 41300 San Jose
de La Rinconada • Sevilla, SPAIN
Telefono: +34 954 150767
Fax: +34 955 266494

info@genovalab.com
www.genova-europe.com

Nova Star Mid Range DNA Ladder

Reference: AB15015



2 of 2

BIBLIOGRAPHY:

Chien A., Edgar D.B., Trela J.M., "Deoxyribonucleic acid polymerase from the extreme thermophile *Thermus aquaticus*", *Journal of Bacteriology*, 127(3), 1550-57, 1976.

Lawyer F.C., Stoffel S., Saiki R.K., Myambo K., Drummond R., et al., "Isolation, characterization, and expression in *Escherichia coli* of the DNA polymerase gene from *Thermus aquaticus*", *The Journal of Biological Chemistry*, 264(11), 6427-37, 1989.

Tindall K.R., Kunkel T.A., "Fidelity of DNA synthesis by the *Thermus aquaticus* DNA polymerase", *Biochemistry*, 27(16), 6008-13, 1988.

Innis M.A., Myambo K.B., Gelfand D.H., Brow M.A., "DNA sequencing with *Thermus aquaticus* DNA polymerase and direct sequencing of polymerase chain reaction-amplified DNA", *Proceedings of the National Academy of Sciences of the United States of America*, 85(24), 9436-40, 1988.

Lo Y.M., Mehal W.Z., Fleming K.A., "Rapid production of vector-free biotinylated probes using the polymerase chain reaction", *Nucleic Acids Research*, 16(17), 8719, 1988.

Erich H.A., (ed.) 1988, "PCR technology: principles and applications for DNA amplification", Stockton Press, New York.

F01IT04_AB15015_V1R1012_EN_Nova_Star_Mid_Range_DNA_Ladder



Catalog number



Batch code



Research use only



Temperature limitation



Expiration date



Manufacturer



See instruction for use



Gennova Scientific, S.L.
C/ Johann Gutenberg, 4F. Pol. Ind.
El Cafamo I • 41300 San Jose
de La Rinconada • Sevilla, SPAIN
Telefono: +34 954 150767
Fax: +34 955 266494

info@gennovalab.com
www.gennova-europe.com